

# Semi-Volatile Organics Analyses



## Drinking Water

EPA Method	Description	Hold Time	Container/Preserve
504.1	EDB and DBCP	14 Day	2-40 ml Vials, Na <sub>2</sub> S <sub>2</sub> O <sub>3</sub>
505	Pesticides and PCB's	14 Day, 24 Hour	2-40 ml Vials, Na <sub>2</sub> S <sub>2</sub> O <sub>3</sub>
515.3	Chlorinated Herbicides	14, 28 Day	2-40 ml Vials, Na <sub>2</sub> S <sub>2</sub> O <sub>3</sub>
525.2	Organic Compounds (Misc. SOCs) <i>chlorinated source</i>	14, 30 Day	2-1L Glass, Na <sub>2</sub> SO <sub>3</sub> + HCl in Field
525.2	Organic Compounds (Misc. SOCs)	14, 30 Day	2-1L Glass, HCl
531.1	Carbamates	28 Day	2-40 ml Vials, Na <sub>2</sub> S <sub>2</sub> O <sub>3</sub> & MCAA
547	Glyphosate	14 Day	2-40 ml Vials, Na <sub>2</sub> S <sub>2</sub> O <sub>3</sub>
548.1	Endothall	7, 14 Day	2-40 ml Vials, Na <sub>2</sub> S <sub>2</sub> O <sub>3</sub>
549.1	Diquat	7, 21 Day	1-500 ml Amber Plastic, Na <sub>2</sub> S <sub>2</sub> O <sub>3</sub> + H <sub>2</sub> SO <sub>4</sub> in field
549.1	Diquat <i>chlorinated source</i>	7, 21 Day	1-500 ml Amber Plastic, Na <sub>2</sub> S <sub>2</sub> O <sub>3</sub>
1613	Dioxin	6 Month	2-1L Glass, None

For further information on LEGEND's analytical capabilities please [contact us](#).